# Safety assessment

## ChroMedium Coliform (Chromogenic Coliform Agar, CCA)

According to Regulation (EC) No. 1907/2006
Issue date: 2013-06-25, Replaces version: xxx-xx-xx

### Article number

TN 1064

### Trade name

ChroMedium Coliform (Chromogenic Coliform Agar, CCA)

### Intended use

Chromogenic selective agar to detect *Escherichia coli* and coliform bacteria from water with low accompanying flora (acc. to the draft standard DIN EN ISO 9308-1:2012) and from processed food.

### Name of supplier

sifin diagnostics gmbh
Berliner Allee 317-321
13088 Berlin, Germany

### Contact for technical information

Phone: +49 30 927 030-0, Head of Department Culture Media
Telefax: +49 30 927 030-30
E-mail: info@sifin.de

### E-Mail (competent person)

msds@conceptec.de

### Emergency telephone number

Phone: +49 30 927 030-0, Safety Officer for Medical Devices
Mo. - Th. 7.30 a.m. - 4.15 p.m., Fr. 7.30 a.m. - 3.00 p.m.

### Classification of the mixture

Classification according to Regulation (EC) 1272/2008 (CLP): This mixture is classified as not hazardous according to Regulation (EC) 1272/2008 (CLP).
Classification according to Directive 1999/45/EC (DPD): This mixture is classified as not hazardous according to Directive 1999/45/EC.

### Label elements

Labelling according to Regulation (EC) 1272/2008 [CLP]: Not applicable.
Labelling according to Directive 1999/45/EC (DPD): Not applicable.

### Other hazards

None known. Minimum standards for protection measures (TRGS 500) as common in the chemical industry should be followed.

**According to Article 31 of the REACH Regulation, drawing up a safety data sheet for this product is unnecessary, as the product is neither classified as a hazardous material nor does it contain hazardous ingredients or materials with Community (EC) workplace exposure limits or substances of very high concern (SVHC) above their particular legal nominal limits. Under REACH Regulation (EC) no. 1907/2006 (REACH), no safety data sheet is therefore required and also not available in this case.**



Manufacturer: sifin diagnostics gmbh | Berliner Allee 317-321 | 13088 Berlin | Germany | T: +49 30 927 030-0 | F: +49 30 927 030-30 | info@sifin.de | www.sifin.de