Safety assessment
Casein Soya Bean Digest Agar (CASO Agar) acc. to harm. EP/USP/JP

According to Regulation (EC) No. 1907/2006
Issue date: 2020-04-01, Replaces version: 2014-10-01

Article number
TN1031, TN1031-01

Trade name
Casein Soya Bean Digest Agar (CASO Agar) acc. to harm. EP/USP/JP.

Intended use
Inibitor-free and indicator-free culture medium. Culture medium for testing non-steril products. The medium fulfils the requirements of the harmonised methods according to EP/USP/JP. No IVD - The laboratory diagnostic examination of human samples has not been evaluated.

Name of supplier
sifin diagnostics gmbh
Berliner Allee 317-321
13088 Berlin, Germany

Contact for technical information
Phone: +49 30 700 144-0, Head of Department Culture Media
Telefax: +49 30 700 144-30
E-mail: info@sifin.de

E-Mail (competent person)
msds@sifin.de

Emergency telephone number
Phone: +49 30 700 144-0, Safety Officer for Medical Devices
Mo. - Th. 7.30 a.m. - 4.15 p.m., Fr. 7.30 a.m. - 3.00 p.m.

Classification of the mixture
Classification according to Regulation (EC) 1272/2008 (CLP): This mixture is classified as not hazardous according to Regulation (EC) 1272/2008 (CLP).
Classification according to Directive 1999/45/EC (DPD): This mixture is classified as not hazardous according to Directive 1999/45/EC.

Label elements
Labelling according to Regulation (EC) 1272/2008 [CLP]: Not applicable.
Labelling according to Directive 1999/45/EC (DPD): Not applicable.

Other hazards
None known. Minimum standards for protection measures (TRGS 500) as common in the chemical industry should be followed.

According to Article 31 of the REACH Regulation, drawing up a safety data sheet for this product is unnecessary, as the product is neither classified as a hazardous material nor does it contain hazardous ingredients or materials with Community (EC) workplace exposure limits or substances of very high concern (SVHC) above their particular legal nominal limits. Under REACH Regulation (EC) no. 1907/2006 (REACH), no safety data sheet is therefore required and also not available in this case.

Manufacturer: sifin diagnostics gmbh | Berliner Allee 317-321 | 13088 Berlin | Germany | T: +49 30 700 144-0 | F: +49 30 700 144-30 | info@sifin.de | www.sifin.de